

Certificate of Analysis

Issue Date: 24.10.2025

STD-No. : A01269-8
Product name Caffeine
Concentration : 300 µg/ml
Solvent : Water

Volume ≥ 1,2 ml
Lot-No. : 54327-005
Expiry date : 04/2027¹
Storage : 20°C and dark

only for information purposes

Component	Conc. µg/ml	Purity	CAS	Formula	M [g/Mol]
1 Caffeine	300 ±3,5 % ²	99,9 % ³	58-08-2	C8H10N4O2	194,19

This Reference Material was processed under the producers ISO 9001:2015 registered quality system.

This Reference Material is intended as working reference sample for identification of the contained compounds and their quantification in methods of analysis in residue or environmental analysis

Traceability to the International System of Units (SI) has been established through an unbroken chain of comparisons, each having stated uncertainties. Comparisons are based on convenient physical or chemical measurements, including gravimetric or volumetric dilution. The balances used for these measurements are calibrated by an accredited calibration service.

The homogeneity is determined according to an in-house procedure. There is no minimum sub-sample required.

Sample aliquots for analysis should be withdrawn at roomtemperature and should be processed without delay for the certified values to persist.

¹ The certification of this reference material is valid within the stated uncertainty until the above specified expiration date assumed the reference material is stored in the originally closed flask in accordance with the instructions given in this certificate. Once the packing has been opened, in the event of improper use no warranty can be given for the certificated values until the expiry date has been reached. The long term stability may be monitored over the lifetime of the certification according to an in-house procedure. If substantive changes are determined that effect the certification before the expiration of this certificate, Campro Scientific GmbH will notify the purchaser.

² Expanded Uncertainty according to EURACHEM / CITAC „Quantifying Uncertainty in Analytical Measurement“ with coverage factor k=2 for a confidence level of 95 %.

³ Stated purity of the neat material is considered in the production of the solution.

This document was produced electronically and is valid without a signature.
Transferred from the original certificates by
Campro Scientific GmbH

