

# Declaration of Conformity

**Producer** KNAUER Wissenschaftliche Geräte GmbH  
Hegauer Weg 38  
14163 Berlin

**Model/Type** Conductivity Monitor CM 2.1S Product no. EDGXXYYZZ  
pH Monitor pH 2.1S X, Z=0-9; Y=A-Z

**Reference** The product complies with the following standards:

---

**EMC** EMC Directive 2014/30/EU  
IEC 61000-3-2:2014  
IEC 61326-1:2012

**Disposal** RoHS Directive 2011/65/EU (2015/863/EU included)  
WEEE Directive 2012/19/EU


**Safety** Low Voltage Directive 2014/35/EU  
IEC 61010-1:2010  
IEC 61010-2-081:2001 + A1



---

The product was tested in a typical configuration. The mark of conformity has been applied to the rear panel. This declaration of conformity applies to above-listed product placed on the EU market after:

**Date** Berlin, 2023-02-13



Alexandra Knauer (CEO and owner)

# Declaration of Conformity for the UK

**Producer** KNAUER Wissenschaftliche Geräte GmbH  
Hegauer Weg 38  
14163 Berlin

**Model/Type** Conductivity Monitor CM 2.1S Product no. EDGXXYYZZ  
pH Monitor pH 2.1S X, Z=0-9; Y=A-Z

**Reference** The products comply with the following standards:

---

**EMC** Electromagnetic Compatibility Regulations 2016  
EN 61000-3-2:2014  
EN 61000-3-3:2013  
EN 55016-2-1:2014+A1:2017  
EN 55016-2-2:2011  
EN IEC 61326-1:2013

**Disposal** RoHS Directive 2011/65/EU (2015/863/EU included)  
WEEE Directive 2012/19/EU  
REACH Regulation (EC) No. 1907/2006 (contains no SVHC above 0,1% (w/w))


**Safety** Electrical Equipment (Safety) Regulations 2016  
BS EN 61010-1:2010  
BS EN 61010-2-081:2015



---

The products were tested in a typical configuration. The UKCA-mark of conformity has been applied to the rear panel. This declaration of conformity applies to above-listed products placed on the UKCA market after:

**Date** Berlin, 2023-02-13



Alexandra Knauer (CEO and owner)