



CERTIFICATE OF ANALYSIS

PRODUCT: 2000 mOsm/Kg H₂O Osmolality Standard
PRODUCT NO: Y1248
LOT NO: Y124819J1
DATE OF TEST: 23rd September 2019
EXPIRY DATE: 28th September 2022
SPECIFICATION: 1980 – 2020 mOsm/Kg H₂O
MEAN: 1982 mOsm/Kg H₂O

TEST METHOD:

The result reported above was determined by analysis of a sample of this lot taken at time of manufacture. The osmolality was determined using a high performance calibrated Osmometer.

Standard Osmolality solutions used for Osmometer calibration have traceability to NIST (SRM) 919b Sodium Chloride via an unbroken chain of comparisons.

This certificate relates solely to the lot number given above.

The uncertainty of measurement has been calculated not to exceed $\pm 1.2\%$ Osmolality value at 95% confidence level, i.e. coverage factor $k=2$.

This certificate is electronically generated and does not require a signature.

Date: 23rd September 2019

This certificate must be reproduced except in full.