



CERTIFICATE OF ANALYSIS

PRODUCT: 300 mOsm/Kg H₂O Osmolality Standard

PRODUCT NO: Y1240

LOT NO: OSM30022C1

DATE OF TEST: 28th March 2022

EXPIRY DATE: 28th March 2025

SPECIFICATION: 298-302 mOsm/Kg H₂O

MEAN: 300 mOsm/Kg H₂O

TEST METHOD:

The result reported above was determined by analysis of a sample of this lot taken at time of manufacture. The osmolality was determined using a high performance calibrated Osmometer.

Standard Osmolality solutions used for Osmometer calibration have traceability to NIST (SRM) 919b Sodium Chloride via an unbroken chain of comparisons.

This certificate relates solely to the lot number given above.

The uncertainty of measurement has been calculated not to exceed $\pm 1.2\%$ Osmolality value at 95% confidence level, i.e. coverage factor $k=2$.

This certificate is electronically generated and does not require a signature.

Date: 28th April 2022

This certificate must be reproduced except in full.